Manual Supplement

Manual Title:	6020 Series Users	Supplement Issue:	**1**
Part Number:	3720771	Issue Date:	3/14
Print Date:	April 2013	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2014 Fluke Corporation. All rights reserved.

FLUKE®
Calibration

Change #1

On page 12, following the **Environmental Conditions** section, add:

Note

The Product may exhibit some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m up to 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.