Manual Supplement

Manual Title:	7005/7015 Users	Supplement Issue:	**3**
Part Number:	3729297	Issue Date:	10/17
Print Date:	September 2014	Page Count:	1
Revision/Date:	740403		

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2014-2017 Fluke Corporation. All rights reserved.

FLUKE®
Calibration

Change #1

On page 8, after the last bullet in the **Environmental Conditions** section, add:

Note

The Product can show some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m to a maximum of 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.

Change #2, 434

On page 7, under **Specifcations**, replace the **Range** with:

Range 0 °C to 50 °C (32 °F to 122 °F)

Change #3, 525

On page 8, under the ***Environmental Conditions***, replace the 1st and 2nd bullets with:

- ambient temperature range: 5 °C to 35 °C (41 °F to 95 °F)
- ambient relative humidity: maximum 80 % for temperatures <31 °C decreasing linearly to 50 % at 35 °C