Manual Supplement

Manual Title:	753/754 Calibration	Supplement Issue:	**4**
Part Number:	Web Only	Issue Date:	9/16
Print Date:	November 2011	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2012-2016 Fluke Corporation. All rights reserved.

FLUKE®

Change #1, 61489

On page 8, in the ***Temperature, Thermocouples*** Table:

Change: Normal Mode Rejection: 40 dB at 50 Hz or 60 Hz nominal

To: Normal Mode Rejection: 65 dB at 50 Hz or 60 Hz nominal

Change #2, 64190, 64295, 544

On page 3, remove the ℂ from the symbols table, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
	Consult user documentation.

Replace the  row with:

	Conforms to relevant Australian Safety and EMC standards.

On page 4, under ***General Specifications***, replace **Power** with:

Power ..... Internal battery pack: Lithium Ion, 7.2 V dc, 30 Wh
External Power Supply: 15 Vdc, 2 A

On page 4, replace the ***Standards and Agency Approval Specifications*** with:

Standards and Agency Approval Specifications

Intrusion Protection ..... IEC 60529-1, IP 52

Safety

General ..... IEC 61010-1, Pollution degree 2

Measurement ..... IEC 61010-30: CAT II 300 V

Electromagnetic Compatibility (EMC)

International ..... IEC 61326-1: Basic Electromagnetic Environment
CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.

Caution: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

Korea (KCC) ..... Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC) ..... 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.

Change #3, 315, 563

On page 5, add the following to the Symbols table:

 Li-ion	This product contains a Lithium-ion battery. Do not mix with solid waste stream. Spent batteries should be disposed of by a qualified recycler or hazardous materials handler per local regulations. Contact your authorized Fluke Service Center for recycling information.

On page 4, under ***Environmental Specifications***, replace **Operating Temperature with Battery,** and add:

Operating Temperature with Battery ..... 0 °C to 40 °C

Battery

Charging Temperature..... 0 °C to 40 °C

Charging Supply ..... Input 100 V ac to 240 V ac 50/60 Hz, ±10 %

Output 15 V dc, 2 A, LPS