Manual Supplement

Manual Title:	7900 Users Guide	Supplement Issue:	**1**
Part Number:	3720895	Issue Date:	3/14
Print Date:		Page Count:	1
Revision/Date:	710801		

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2014 Fluke Corporation. All rights reserved.

FLUKE®
Calibration

Change #1

On page 10, after the last bullet in the **Environmental Conditions** section, add:

Note

The Product can show some control sensitivity to moderate or severe electromagnetic fields or conducted interference of certain frequencies. In the presence of radiated EM disturbances, with frequencies of 250 MHz to 400 MHz and with amplitude >1 V/m to a maximum of 3 V/m, add 0.0025 °C to the stability specification. Stability is not guaranteed if amplitude is >3 V/m. When subject to conducted disturbances of 8 MHz to 80 MHz, and amplitude >3 V, add 0.005 °C to the stability specification.