Supplement

Manual Title:	a3004 FC Safety Information		
Part Number:	4602181	Supplement Issue:	**1**
Print Date:	March 2015	Issue Date:	8/16
Revision/Date:		Page Count:	2

This supplement contains information necessary to ensure the accuracy of the above manual.

© 2016 Fluke Corporation. All rights reserved.

FLUKE®

Change #1, 504

On the second page, replace the ***Safety Specifications*** with:

Safety Specifications

Maximum Voltage between Terminals and Earth Ground: 30 V

Range ..... 120 m Adc (with Fluke i20 Current Probe)

Battery..... Two AA, NEDA 15 A, IEC LR6

Temperature

Operating Temperature..... -10 °C to +50 °C

Storage Temperature..... -40 °C to +60 °C

Operating Relative Humidity. 90 % at 35 °C
75 % at 40 °C
45 % at 50 °C

Altitude

Operating ..... 2000 m

Storage..... 12 000 m

Ingress Protection..... IEC 60529: IP42 (non-operating)

Safety

IEC 61010-1..... Pollution Degree 2

Electromagnetic Compatibility (EMC)

International ..... IEC 61326-1: Portable Electromagnetic Environment;
IEC 61326-2-1

CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.

Caution: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

Korea (KCC)..... Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC) ..... 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.